

This product contains hydrogen sulfide and nitrogen, substances subject to the Pennsylvania Worker and Community Right-To-Know Act.

PRODUCT IDENTITY

LABEL IDENTITY - MSA P/N 711060 Calibration Check Gas, 10 ppm Hydrogen Sulfide in Nitrogen
CHEMICAL NAME - Hydrogen Sulfide, Nitrogen Mixture
ADDITIONAL IDENTITIES - MSA P/N 711060 Calibration Gas
FORMULA - H₂S in N₂

APPLICABLE CHEMICAL CONTENTS

	<u>%</u>	<u>TWA</u>
Hydrogen Sulfide (CAS 7783-06-4) STEL 15 ppm (ACGIH 1995-96)	0.001	10 ppm
Nitrogen	Balance	None

NOTE: Gas Under Pressure, 500 PSIG at 70°F, Approx. 34 Liters Gas at Atmospheric Pressure

PHYSICAL AND CHEMICAL PROPERTIES

APPEARANCE AND ODOR - Colorless, Malodorous Gas
BOILING POINT - N/A
VAPOR PRESSURE - N/A
VAPOR DENSITY (AIR = 1) - < 1
SOLUBILITY IN WATER - H₂S - 437 cm³/100 ml (0°C)
Nitrogen - 2.3 cm³/100 ml (0°C)
SPECIFIC GRAVITY (H₂O = 1) - N/A
PERCENT VOLATILE BY VOLUME - N/A

N/A - Not Applicable

PHYSICAL HAZARD INFORMATION

PHYSICAL HAZARD - Compressed Gas, 500 PSIG at 70°F
CONDITIONS OR MATERIALS TO AVOID - None
FLASH POINT - N/A
LEL - N/A
UEL - N/A
EXTINGUISHING MEDIA - This Calibration Gas Mixture Is Not Flammable
SPECIAL FIRE FIGHTING PROCEDURES - See Next Item
UNUSUAL FIRE AND EXPLOSION HAZARDS - Gas Under Pressure, 500 PSIG at 70°F. Do Not Exceed 120°F.

HEALTH HAZARDS

HEALTH HAZARDS - LCLO 800 ppm/5 minutes (H₂S). H₂S may cause pulmonary edema, headache, dizziness, bronchitis, and respiratory paralysis. Nitrogen is an asphyxiant.

SIGNS AND SYMPTOMS OF EXPOSURE - H₂S is an irritant of the eyes and respiratory tract.

PRIMARY ROUTES OF ENTRY - Inhalation, Eyes

TARGET ORGANS - H₂S is an irritant of the eyes and respiratory tract.

MEDICAL CONDITIONS GENERALLY RECOGNIZED AS BEING AGGRAVATED BY EXPOSURE - No Information

EXPOSURE LIMITS - ACGIH 1995-96: H₂S 10 ppm, STEL 15 ppm

CARCINOGENICITY DATA - Component gases are not listed by NIOSH RTECS, OSHA, NTP or IARC.

EMERGENCY AND FIRST AID PROCEDURES - Remove From Exposure. If breathing has stopped, give artificial respiration. Consult Physician Immediately.

SAFE HANDLING AND USE

HYGIENIC PRACTICES - Avoid Breathing Gas

PROTECTIVE MEASURES DURING REPAIR AND MAINTENANCE OF CONTAMINATED EQUIPMENT - N/A

PROCEDURES FOR SPILL OR LEAK CLEANUP - Ventilate Area. Avoid Breathing Gas.

WASTE DISPOSAL - Do not puncture or incinerate cylinder. Before discarding cylinder, slowly release contents to a safe exhaust.

STORAGE - Store in a cool, dry, well-ventilated area. Do not exceed 120°F.

CONTROL MEASURES

PERSONAL PROTECTIVE EQUIPMENT - Due to the limited amount of gas in the cylinder, and the low release rate employed in instrument calibration, respiratory protection is not indicated under conditions of intended use.

ENGINEERING CONTROLS - Mechanical ventilation is suitable.

WORK PRACTICES - Avoid breathing gas. Use in well-ventilated areas. Follow the calibration procedure detailed in the MSA instruction manual provided with the instrument under calibration.

DATE OF PREPARATION - Rev. 4, June 1999

WARNING: This is a hazardous chemical product. By following the directions and warnings provided with this product, the hazards associated with the use of this product can be greatly reduced but never entirely eliminated. Mine Safety Appliances Company makes no warranties, expressed or implied, with respect to this product and EXPRESSLY DISCLAIMS THE WARRANTY OF MERCHANTABILITY AND ANY WARRANTY OF FITNESS FOR A PARTICULAR PURPOSE. Users assume all risks in handling, using or storing this product.